

ARKANSAS STATE HIGHWAY & TRANSPORTATION DEPARTMENT
EQUIPMENT AND PROCUREMENT OFFICE
LITTLE ROCK, ARKANSAS

March 21, 2007

NOTICE TO BIDDERS
ADDENDUM

BID NO. M-07-106P

Wavelength-Dispersive X-Ray Fluorescence (XRF) Spectrometer and Accessories
OPENING DATE: April 3, 2007

Specifications for the above bid has been changed to read as follows:

Specimen Chamber and Introduction System – 1st paragraph should have read as follows:

The unit will be equipped with a specimen changer that will accept solid or liquid samples or both. It shall allow for different sizes and shapes of specimens at the same time without disassembly of the changer. It shall allow for automatic loading of the specimens. It shall be capable of differentiating between liquid and solid samples. Safeguards shall be employed to allow for proper analysis and to prevent the inadvertent introduction of contamination into the spectrometer optical path. A particulate collection system to remove any loose matter from the loading chamber must be provided on systems that place samples directly in the spectrometer.

Warranty and Support – The following paragraph needs to be added.

A five-year unconditional full warranty of the optical system including cleaning and replacement of collimators, crystals, window, seals, pumps, x-ray tubes and other related items is to be provided in the event of failure of the Specimen Chamber and Introduction System safeguards.

A SIGNED COPY OF THIS ADDENDUM MUST BE SUBMITTED
WITH A SIGNED COPY OF BID INVITATION.

Danny Keene
Division Head
Equipment and Procurement

Company Name: _____

Vendor Signature: _____

Signature must be original (not photocopied) and in ink.